

Hazard -

Hazard - EMF Exposure From:	Persons at Particular Risk	ADDITIONAL CONTROLS: Persons at Particular Risk
Indirect Effects: Static magnetic field could interfere with Passive and Active Medical Implanted Devices (AMIDs)	Consider: As potential for interference with Passive and Active Medical Implanted Devices (AMIDs) consider: Persons working in close proximity to machine eg loading samples and working with probes at base	As General Controls above PLUS RESTRICT persons fitted with medical implanted devices from: Loading samples Working at base



